

Manufacturing Quality

ISO 9001:2015

ISO 13485:2016

Certificate of Analysis

Onco-Ref™ Genomic DNA Reference Standard HCT116 WT


Product Information

Catalog Number:	ASO-6052-1
Lot Number:	GHCT190610
Expiration Date:	2021-06-09
Quantity/Volume/Size:	5.0 µg / 100 µl / 1 vial
Cell:	HCT116
Storage Temperature:	-20°C
Shipping Temperature:	Ambient
Instructions for Use:	Refer to product Datasheet (Available www.accuref.com)
Intended Use:	Research Use Only. Not Intended for human or animal diagnostic or therapeutic use.

Quality Control

Test	Method	Acceptance Criteria	Result
Quality	Gel Electrophoresis	No Smear	PASS
Quantification	Qubit 4.0 Fluorometer	50 -60 ng /µl	PASS
Genotype WT	Sanger Sequencing/ NGS	No New Mutation Detected	PASS

AccuRef Diagnostics certifies that this standard meets the specifications stated in this certificate and warrants this product to meet the stated acceptance criteria through the expiration date when stored unopened as recommended.

	2019-06-19
Quality Assurance	Date

COA-ASO-6052-1-V2